

Numerical simulation of experiments for X-ray diagnostics of pulsed plasma

Skobliakov A V[®], Kolesnikov D S, Kantsyrev A V and Golubev A A

National Research Center “Kurchatov Institute”, Kurchatov Square 1, Moscow 123182, Russia

[®] dinAlt220@yandex.ru

Grazing incidence spectrographs are used to record X-ray spectra of pulsed plasma radiation (for instance, in experiments with high-power lasers or Z-pinches). The main problems that prevent reliable qualitative and quantitative reconstruction of the initial spectra from the data read in the registration plane of the spectrograph (detector plane) are the superposition of signals from different diffraction orders and the complex form of the instrumental function of the spectrograph, which depends on the wavelength of the detected radiation. This problem does not have a trivial analytical solution. The construction of a full-scale Monte Carlo model of an X-ray spectrograph makes it possible to calculate the instrumental function of the device used to reconstruct the initial X-ray spectrum. The simulation takes into account the probabilities of reflection in the diffraction orders after interaction with the diffraction grating, taking into account the surface profile of the diffraction grating [1–3].

- [1] Botten L C Cadilhac M *et al* 1980
- [2] Maystre D Popov E 2014
- [3] Goray L I Schmidt G 2014